

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 27 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130124</u>
A	Name	<u>NICHOLAS YLTIEMI</u>	ID# <u>3429</u> Date <u>5/20/09</u>
	Agency	<u>MEILAKATLA POLICE DEPT.</u>	Phone # <u>886-4011</u>
	Instrument Location	<u>MEILAKATLA POLICE DEPT.</u>	
B	Alco S/N	<u>X172942</u>	Target Value <u>.084</u> High Pressure <u>1300 psi</u>
	Alco Test Values	<u>.087</u> 1 st Alco	<u>.087</u> 2 nd Alco
	Signature	<u>Nicholas Yltemi</u>	
(OVER)			

BMB
5/28/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz

Scientific Director


State Breath Alcohol Program

Subscribed and sworn before me this 18th day of June 2009.

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130124 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

MAY 20, 2009

OPERATOR'S NAME:

VLINTINI/NICHOLAS/L
OPERATOR'S NUMBER: 3429

SUBJECT'S LAST NAME:
UER

SUBJECT'S FIRST NAME/MI:
OF CAL

O.L. #: 0000

DEPT/AGENCY: MET1

CASE/REPORT: 0000

TEST TYPE: U

ALCO TARGET VALUE: .004 ✓

ALCO S/N: X172942 ✓

BREATH ANALYSIS

.004 ADJUSTED FOR 30.45 in ✓
ALCO TARGET .005 20:14
BLANK TEST .000 20:16
INTERNAL STANDARD VERIFIED 20:16
ALCO TO 30.45 in .007 ✓ 20:16
BLANK TEST .000 20:17
SUBJECT SAMPLE .000 20:18
BLANK TEST .000 20:18
ALCO TO 30.45 in .007 ✓ 20:19
BLANK TEST .000 20:19

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

MAY 20, 2009 ✓
TIME 20:20

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.45 in
FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@AB CDEF G
HIJKLMN O PQRSTU VWXYZ[\]^_`abcde fghijklmnop
q rstuvwxyz{|}~■